



# Scantibodies

## Whole PTH (1-84) Specific ImmunoChemiluminoMetric Assay (ICMA) (Coated Plate-Technology)

(Part Number: 3KG002)

For the quantitative determination  
of human Whole PTH

**\*\*For *In Vitro* Diagnostic Use Only\*\***

Store at 2 - 8° C

### INTENDED USE

This kit has been designed for the quantitative determination of human whole parathyroid hormone (PTH) without cross-reaction to PTH (7-84) fragment in plasma samples.<sup>32,33,34,35</sup>

### PHYSIOLOGY

The Whole PTH peptide (1-84) is secreted by parathyroid glands under the regulation of the extracellular concentration of ionized calcium, vitamin D and magnesium. PTH acts with respect to calcium on the kidney and the skeleton<sup>4,5</sup>. PTH binds to receptors, which stimulate adenylate-cyclase to produce cyclic adenosine monophosphate (cAMP) from adenosine triphosphate (ATP)<sup>10,13</sup>. The biological activity of PTH resides in the first 3 amino acids of the N-terminal portion of the molecule. PTH is metabolized either intraglandular or in the peripheral organs into fragments. Circulatory PTH are immunologically heterogeneous<sup>7,6,12,13,18,19</sup>. A recent study of circulation immunoreactive PTH showed that significant amounts of a large carboxyl-terminal PTH fragment, PTH (7-84), presented in blood samples from uremic patients<sup>31</sup>. Biologically inactive fragments with molecular weights of 4000 - 7000 Daltons circulate with a half-life of 30 minutes in healthy persons<sup>4,5</sup>.

cAMP or other PTH dependent processed metabolites (e.g. hypophosphatemia) stimulate the renal hydroxylation of 25-(OH) vitamin D to 1,25-(OH)<sub>2</sub> vitamin D. This vitamin D metabolite stimulates calcium absorption by the small intestine. Severe vitamin D deficiency results in an enhanced secretion of PTH compared to the secretion of calcium. Hypomagnesemia in the primary stage stimulates hypocalcemia. Severe hypomagnesemia results in the reduced secretion of PTH.

Primary and secondary hyperparathyroidism, kidney insufficiency, malabsorption-syndrome and pseudo-hypoparathyroidism result in elevated concentrations of PTH<sup>14,15,16</sup>. Decreased concentrations of PTH coincide with high doses of vitamin-D, milk-alkali-syndrome, Morbus Boeck, hyperthyreosis, ingestion of thiazide and hypercalcemia of malignancy. PTH concentration is also decreased with absorptive hypercalciuria and hypoparathyroidism.

### PRINCIPLE OF PROCEDURE

Scantibodies 1-84 PTH or Whole PTH Kit is an immunochemiluminometric assay (ICMA) utilizing a polyclonal PTH 1-84 antibody with a tendency to bind in the N-terminal region of PTH 1-84 (Label Antibody), and a polyclonal PTH 1-84 antibody with a tendency to bind in the C-terminal region of PTH 1-84 (Capture Antibody). The use of these antibodies guarantees that only Whole PTH (CAP) is detected. The Label Antibody is labeled with isoluminol. The Capture Antibody is fixed to the wells of the microtiter plate. 1-84 PTH or Whole PTH (CAP) in patient samples is bound both to the wells of the microtiter plate strips and to the Label Antibody. After incubation simple wash steps reduce the non-specific binding (NSB) to a minimum for increased precision at the low end of the calibration curve. The concentration of Whole PTH (CAP) is directly proportional to the photons emitted from the wells upon addition of triggering reagents. The concentration of PTH in unknown patient samples and controls is determined by interpolation using a calibration curve.<sup>30</sup>

### REAGENTS

The Scantibodies Whole PTH Kit contains sufficient reagents for 96 single determinations. The kit is stable at 2° - 8° C until the stated expiration date.

### PTH CALIBRATORS

One set of calibrators consists of seven vials containing lyophilized human serum with nominal Whole PTH (CAP) concentrations. The lyophilized calibrators are prepared in stabilized human serum containing sodium azide 0.1% (w/v). The Whole PTH (CAP) concentrations are declared on the vial label.

### PTH CONTROLS

One set of controls consists of two vials containing Whole PTH (CAP) in lyophilized human serum with 0.1% (w/v) sodium azide. The concentration ranges of Whole PTH (CAP) are declared on the vial labels.

### WHOLE PTH TRACER FOR ICMA

The tracer consists of one bottle of goat anti N-terminal PTH antibodies labeled with isoluminol and dissolved in 10 ml phosphate buffered saline with a non-azide, non-mercury preservative 0.2% (v/v) and

Scantibodies Laboratory, Inc.

This diagnostic kit complies with IVDD 98/79/CE

protein stabilizers.

### **PTH ANTIBODY COATED PLATE**

One plate contains 12 strips (96 wells) in a frame plus desiccant. The strip wells are coated with the capture antibody. The desiccant contains silica.

### **WASH CONCENTRATE**

One bottle contains 30 ml of a 10 fold concentrate of phosphate buffered saline with sodium azide 0.6% (w/v) and detergent.

#### **TRIGGER 1**

One bottle contains 28 ml of ready to use reagent [Caution! Contains 1M sodium hydroxide]

#### **TRIGGER 2**

One bottle contains 28 ml of ready to use reagent

## **PREPARATION AND STORAGE OF REAGENTS**

### **PTH CALIBRATORS**

The Scantibodies Laboratory, Inc. Whole PTH ICMA Diagnostic Kit contains PTH standards prepared analytically on a mass basis from purified synthetic Whole PTH (1-84). These standards are further evaluated against lyophilized "primary standards" which are stored at -70° C to maintain potency.

Reconstitute the zero calibrator with 5 ml of distilled or deionized water. Reconstitute the remaining calibrators with 2 ml of distilled or deionized water. After addition of the water, mix each vial thoroughly but gently. Use the reconstituted calibrators within 1 hour. Store the unused portion of the calibrators below -20° C until the stated expiration date. Do not store the calibrators at room temperature for more than one hour at any given time. Do not thaw any calibrator vial more than two times.

### **PTH CONTROLS**

Reconstitute the vials of controls with 2 ml of distilled or deionized water. After addition of the water, mix each vial thoroughly but gently. Use the reconstituted controls within 1 hour. Store the unused portion of the controls below -20° C until the stated expiration date. Do not store the controls at room temperature for more than one hour at any given time. Do not thaw any control vial more than two times.

### **WHOLE PTH ICMA TRACER**

The tracer is ready to use. Store the tracer at 2° - 8° C until the stated expiration date.

### **PTH (39-84) ANTIBODY COATED PLATE**

The antibody coated plate is ready to use. Store the plate at 2° - 8° C until the stated expiration date. Allow the plate to equilibrate to ambient temperature prior to opening package. Reseal the

package immediately after removing the required number of strips.

### **WASH CONCENTRATE (10X)**

Mix the contents of the wash concentrate thoroughly with 270 ml of distilled or deionized water (1:10) to make the wash solution. Store the wash solution at room temperature (18° - 25° C) until the stated expiration date. **Note:** On occasion, the wash concentrate may partially crystallize depending on individual storage conditions. In this case, additional mixing may be required in order to achieve complete dissolution of the reagent.

#### **TRIGGER 1**

The reagent is ready to use. Store at 2° - 8° C until the stated expiration date.

#### **TRIGGER 2**

The reagent is ready to use. Store at 2° - 8° C until the stated expiration date.

## **WARNINGS AND PRECAUTIONS FOR USERS**

### **Use of the Assay**

The reagents are for in vitro diagnostic use only.

### **Human Serum Caution**

The human serum in this kit has been prepared from human donors and it has been tested by FDA approved immunoassays and found to be non-reactive for Hepatitis B Surface Antigen (HBsAg), Anti HIV I/II and Anti HCV. However, it is recommended to consider the calibrators and controls as a potential biohazard and handle them with the same precautions as applied to any untested patient sample.

### **Sodium Azide (NaN<sub>3</sub>) Warning**

Some reagents in the Scantibodies PTH assay contain sodium azide. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. On disposal flush the drain with a large volume of water to prevent azide build-up. Avoid direct contact with skin and mucous membranes.

### **Corrosive Warning**

Trigger 1 reagent contains 1M sodium hydroxide which is very corrosive. Do not allow to contact skin or splash into eyes. Wear gloves and goggles when handling this reagent. Dispose of any unused reagent in a safe manner according to local regulations.

## SAMPLE PREPARATION AND STORAGE

### Specimen Collection

The determination of human Whole PTH should be made on EDTA-plasma. Two hundred microliters of plasma are required to assay one sample in duplicate. To obtain plasma, collect blood by venipuncture into a tube containing EDTA. Invert the tube gently 5 - 6 times after collection to ensure adequate mixing. Whole blood may be stored refrigerated for up to 48 hours prior to centrifugation. Centrifuge the sample and separate the plasma from the cells. Plasma should be stored at -20° C or lower. Avoid repeated freezing and thawing of plasma. Do not use patient samples which have been frozen and thawed more than two times.

### Dilution of Patient Samples

Dilute plasma samples with PTH concentrations greater than the highest calibrator with Scantibodies PTH Zero Calibrator before assay. The dilution factor is applied to the diluted sample assay result in order to determine the PTH concentration in the undiluted sample.

### Quality Control

Two levels of controls are provided with each assay kit. The values assigned to these controls are printed on the container label. The control value should fall within the specified range when tested in the same manner as the unknowns. Controls should be included in each assay. If the control values do not meet the established range, the assay may be invalid and should be repeated.

## ASSAY PROCEDURE

### Materials Provided

The Scantibodies Whole PTH Kit (Part No. 3KG002) is supplied with the following:

Description	Number
PTH Standards 3CA650, 3CB650, 3CC650, 3CD650, 3CE650, 3CF650, 3CG650	7 vials
PTH Controls Part Nos. 3CA651, 3CB651	2 vials
PTH (39-84) Antibody Coated Plate Part No. 3KP004	1 Plate (12 Strips)
Whole PTH Tracer, ICMA Part No. 3KL102	1 vial
Wash Concentrate (10X), ICMA Part No. 3KW002	1 bottle
Trigger 1 Part No. 3KL057	1 bottle
Trigger 2 Part No. 3KL055	1 bottle
Directional Insert Part No. 7KI035	1 insert

### Materials And Equipment Required But Not Provided:

Distilled or deionized water  
Pipettors with disposable tips: 0.05 and 0.1 ml  
Wash station  
VWR orbital shaker Model DS-500 or equivalent  
Microplate luminometer equipped with dual injectors capable of delivering 100  $\mu$ l each.

### Preparation for Assay

For each assay, prepare the following number of strips for double determination:

- 2 strips for calibrators A – G and control I
- 2 strips for control II and samples 1 through 7
- 2 strips for samples 8 through 15, etc.

Place any unused strips back into the resealable foil pouch and store at 2° - 8° C

### Pipetting and Incubation Steps

1. Pipette 100  $\mu$ l of calibrators, controls and samples into the bottom of the corresponding strip wells. (Plate mapping forms are useful for recording sample locations on the plate.)
2. Pipette 100  $\mu$ l of Whole PTH tracer into each well changing pipet tips between each well.
3. Place the plate securely on an orbital shaker and rotate the plate for 2 hours to 2½ hours at 160 to 180 RPM at room temperature (18° – 25° C). There is no need to cover the plate or protect from the light.
4. After the incubation time is completed, wash the plate strips by aspirating the contents of each well completely and dispensing 350 to 400  $\mu$ l of wash solution into each well. Repeat this procedure 3 to 5 times aspirating the wash solution completely after the last washing step. (Slapping the plate on absorbent material after washing is **not** recommended).
5. While the plate is being washed, prime the luminometer with enough of the trigger solutions to fully expel any air or liquid in the system.
6. Set up a standard curve in the instrument using the concentrations listed on the standard vials for the determination of the unknown samples.
7. Set the luminometer to inject 100  $\mu$ l of trigger 1 followed by 100  $\mu$ l of trigger 2 with a 0.5 second delay prior to reading and an integration time of 1.5 seconds. The delay and integration settings are general guidelines. Different instruments may require some adjustments to obtain optimal results.
8. Read the plate within 5 minutes after the washing step is completed.

## PIPETTING GUIDE

Additive to well	Calibrator Wells	Control Wells	Sample Wells
Calibrator	100 $\mu$ l	-----	-----
Control	-----	100 $\mu$ l	-----
Sample	-----	-----	100 $\mu$ l
Tracer	100 $\mu$ l	100 $\mu$ l	100 $\mu$ l

Rotate the uncovered plate at 160 to 180 RPM on an orbital shaker for 2 to 2½ hours.

Wash the strips 3 to 5 times with the diluted wash solution.

Read the strips on a luminometer equipped with dual injectors primed with the two trigger solutions and set to deliver 100  $\mu$ l of trigger 1 followed by 100  $\mu$ l of trigger 2 with a 0.5 second delay prior to reading and a 1.5 second integration time, within 5 minutes after the washing step is completed.

## PROCEDURAL COMMENTS

### Interferences:

Samples containing up to 250 mg/dl triglyceride, 15 mg/dl hemoglobin and 30 mg/dl bilirubin do not exhibit any significant effect on the assay. However, it is strongly recommended that grossly hemolyzed or lipemic samples not be used in this assay.

### Reagents from different lot numbers must not be interchanged.

The patient sample or calibrator should be pipetted carefully into the very bottom of the assay well. The tracer should be pipetted onto the side of the well just above the liquid level.

Accurate and complete reconstitution of the controls and standards is essential to achieving accurate assay results.

The washing step is an important step in the assay procedure. Accurate dispensing of the wash solution and complete aspiration of the tube contents is essential to achieving assay sensitivity, low background and assay precision. Washing the wells 5 times with the provided wash reagent is recommended.

It is recommended that calibrators and patient samples be assayed in duplicate. The average Relative Light Units of each duplicate should then be used for data reduction and the calculation of results.

Avoid sample to sample contamination by using a new pipette tip for each sample.

## CALCULATION OF RESULTS

### Evaluation

1. Calculate the average RLU for each double determination.
2. Draw the calibration curve by plotting the average RLU from each duplicate calibrator level (ordinate) against the respective concentration declared on

the calibrator vial (abscissa) using log-log graph paper. Obtain sample concentrations by interpolation of average sample RLU on the calibration curve.


3. If samples were run with dilution, multiply the diluted sample assay results from the curve by the appropriate dilution factors to obtain the undiluted sample assay results.

## SAMPLE DATA

Wells	RLU	Ave. RLU
0.5 pg/ml	394 380	387
5.7 pg/ml	1085 1015	1050
14.0 pg/ml	2043 2011	2027
32 pg/ml	4352 4335	4344
145 pg/ml	16414 16791	16603
805 pg/ml	96843 84474	90659
2294 pg/ml	242130 236820	239475

**NOTE:** The data presented are for demonstration purposes only and must not be used in place of data generated at the time of the assay.

## REPRESENTATIVE STANDARD CURVE



Automated data reduction can also be used to construct the Scantibodies PTH calibration curve. To program automated data reduction systems or to adapt an existing program consult the data processor manufacturer or the programmer.

## LIMITATIONS OF THE PROCEDURE

For diagnostic purposes PTH values should be used in addition to other diagnostic data and clinical information available to the physician.

The assay procedure must be followed exactly; careful technique must be used to obtain valid results. Any modification of the assay procedure is likely to alter the results.

Grossly hemolyzed, lipemic or icteric samples are likely to give non valid results.

The highest concentration of PTH measurable without sample dilution is the concentration of the highest calibrator. The lowest level measurable is below 1.2 pg/ml.

### EXPECTED VALUES

The normal value range was determined following the NCCLS guidelines (C28-A) using 120 samples from apparently healthy individuals. It is recommended that each laboratory establish its own range of normal values. The values given are only indicative and may vary from other published data.

PATIENT CLASSIFICATION	Whole PTH RANGE (pg/ml)
Normal	5 - 39
Hyperparathyroidism	> 39

### PERFORMANCE CHARACTERISTICS

#### Accuracy, Recovery

Three different samples with known concentrations of PTH were spiked with known amounts of PTH. The % recovery was determined following assay of the spiked samples.

Sample value (pg/ml)	Added PTH (pg/ml)	Measured value (pg/ml)	Expected value (pg/ml)	Recovery (%)
132.41	58.41	150.04	190.82	78.8
456.23	52.15	494.52	508.38	97.3
1213.50	147.14	1432.10	1360.64	105.3

#### Accuracy, Dilution

Patient samples with high concentrations of PTH were diluted with Standard A. The % recovery was determined following assay of the diluted samples.

Sample	Dilution	Measured value (pg/ml)	Expected value (pg/ml)	Recovery %
1	Neat	537.27		
	1:2	283.19	268.64	105.4
	1:4	136.92	134.32	101.9
	1:8	64.14	67.16	95.5
2	Neat	752.98		

Sample	Dilution	Measured value (pg/ml)	Expected value (pg/ml)	Recovery %
	1:2	416.13	376.49	110.5
	1:4	208.07	188.25	105.8
	1:8	104.03	94.12	101.9
3	Neat	944.80		
	1:2	523.39	472.40	110.8
	1:4	261.56	236.20	110.7
4	1:8	122.04	118.10	103.3
	Neat	175.66		
	1:2	98.73	87.83	112.4
5	1:4	48.20	43.92	109.8
	1:8	25.62	21.96	116.7
	Neat	211.51		
6	1:2	108.87	105.76	102.9
	1:4	54.18	52.88	102.5
	1:8	26.22	26.44	99.2
	Neat	186.92		
	1:2	98.85	93.46	105.8
	1:4	49.22	46.73	105.3
	1:8	26.57	23.37	113.7
				<b>106.3</b>

#### High Dose Hook Response

This high dose hook response of the Scantibodies Laboratory, Inc. Whole PTH (1-84) Specific ICMA Diagnostic Kit was determined as 20,000 pg/ml of Whole PTH (CAP). Samples greater than the highest standard (approximately 2300 pg/ml) and up to 20,000 pg/ml Whole PTH (CAP) will produce RLU values greater than that of the highest standard.

#### Precision

The inter-assay precision was evaluated by performing 20 separate Whole PTH (CAP) assays on three samples in duplicate over a two week period.

Precision (Inter-assay)			
Sample	Mean value (pg/ml)	SD (pg/ml)	% CV
1	68.70	5.51	8.02
2	208.90	13.53	6.47
3	501.90	36.59	7.29

The intra-assay precision was evaluated by performing 20 replicates in the Whole PTH (CAP) assays on three samples.

Precision (Intra-assay)				
Kit Batch	Sample	Mean value (pg/ml)	SD (pg/ml)	% CV
E1	1	60.94	4.28	7.03
	2	203.31	7.14	3.51
	3	477.49	37.31	7.81
E2	1	63.07	2.99	4.75
	2	199.88	8.19	4.10
	3	483.60	29.30	6.06
E3	1	70.44	3.15	4.47
	2	196.85	6.96	3.54

	3	430.38	13.51	3.14
--	---	--------	-------	------

### Sensitivity

The detection limit of the assay is defined as the lowest measurable value distinguishable from calibrator A. The sensitivity was determined by assaying calibrator A 20 times in the same assay. The minimum detectable dose was below 1.2 pg/ml at 2 standard deviations above the first PTH calibrator.

### Specificity



























This Whole PTH (CAP) assay does not show any cross-reaction to PTH (7-84) fragment when the synthetic PTH (7-84) peptide is serially diluted with standard A matrix and assayed.









PTH (7-84) Conc. Sample (pg/ml)	Measured PTH conc. (pg/ml)
2500	undetectable
5000	undetectable
10000	undetectable
20000	undetectable

A high degree of correlation exists between the PTH levels of duplicate samples measured by a commercially available predicate PTH kit and those levels measured by the Scantibodies Laboratory, Inc. (SLI) Whole PTH (1-84) (CAP) Specific ICMA Assay. A correlation coefficient (r) of 0.984 (n=240) was obtained with a slope of 0.976 and intercept of 6.901 where x represents the predicate device data and y represents the SLI Whole PTH ICMA data. Calculations were made with samples ranging from 1.23 – 995.82 pg/mL.

<b>Chemical Characterization:</b>	1) Antibodies coated on to 96 well polystyrene plates packaged with silica desiccant.
	2) Antibody Tracer labeled with Luminol in a phosphate stabilizer with a mercury preservative @ 0.2%.
	3) Calibrators & controls – human serum containing Sodium Azide @ 0.1%
	4) Antibody Trigger containing Sodium Hydroxide @ 0.4%.
	5) Antibody Trigger containing Hydrogen Peroxide @ 0.5%.
	6) Wash Concentrate containing Sodium Azide @ 0.9%.
<b>Hazardous Ingredients:</b>	Sodium Hydroxide @ 0.4% CAS Number: 1310-73-2 Symbols: Very Toxic; Corrosive R-phrases: R26/27/28, R32, R50/53 S-phrases: S28, S45, S53
	Hydrogen Peroxide @ 0.5% CAS Number: 7722-84-1 Symbols: Very Toxic; Corrosive; Oxidizing R-phrases: R20/21/22, R28, R32, R34, R50/53 S-phrases: S1/2, S3, S28, S36/39, S45, S53

	Sodium Azide @ 0.1% CAS Number: 026628-22-8 Symbols: N/A R-phrases: N/A S-phrases: N/A
English	Sodium Azide @ 0.9% CAS Number: 026628-22-8 Symbols: Toxic R-phrases: R25, R31, R52/53 S-phrases: S28, S45, S53, S60, S61

Symbol	Used for	Symbol	Used for
	Do Not Reuse		Use By YYYY-MM-DD or YYYY-MM
	Batch Code		Serial Number
	Date of Manufacture		Sterile
	Sterilized Using Aseptic Processing Technique		Catalog Number
	Caution, Consult Accompanying Documents		Biological Risks
	Manufacturer		Authorized Representative in the European Community
	Contains Sufficient for < n > Tests		For IVD Performance Evaluation Only
	<i>In vitro</i> Diagnostic Medical Device		Upper Limit of Temperature
	Lower Limit of Temperature		Temperature Limitation
	Consult Instructions for Use		Positive Control
	Control		Negative Control
	Antibody Coated Beads		Antibody Coated Tubes
	Radioactive Iodine Tracer		Wash Solution

Symbol	Used for	Symbol	Used for
	Reconstituted with 2 mL Water		Reconstituted with 5 mL Water
	European Conformity Mark		Radioactive
	Toxic		Harmful
	Corrosive		Oxidizing

### PTH LITERATURES

- Berson, S.A., Yalow, R.S., Aurbach, G.D., and Potts Jr., J.T. "Immunoassay of Bovine and Human Parathyroid Hormone." **Proc. National Academy Science, U.S.A.** 49:613-617, 1963.
- Keutmann, H.T., Sauer, M.M., Hendy, G.N., O'Riordan, J.L.H., and Potts Jr., J.T. "Complete Amino Acid Sequence of Human Parathyroid Hormone." **Biochemistry** 17:5723-5729, 1978.
- Raisz, L.G., Yajnik, C.H. Bockman, R.S., and Bower, B.B. "Comparison of Commercially Available Parathyroid Hormone Immunoassay in the Differential Diagnosis of Hypercalcemia Due to Primary Hyperparathyroidism or Malignancy." **Annals International Medicine** 91:739-740, 1979.
- Habener, J.F., and Potts Jr., J.T. "Biosynthesis of Parathyroid Hormone." **New England Journal of Medicine** 299:580-585, and 635-644, 1978.
- Segre, G.V., D'Amour, P.D., Hultman, A., and Potts Jr., J.T. "Effects of Hepatectomy and Nephrectomy Uremia on Metabolism of Parathyroid Hormone in the Rat." **Journal of Clinical Investigation** 67:439-448, 1981.
- Segre, G.V., Perkins, A.S., Witters, L.A., and Potts Jr., J.T. "Metabolism of Parathyroid Hormone by Isolated Kupffer Cells and Hepatocytes." **Journal Clinical Investigations** 67:449-457, 1981.
- Segre, G.V., Habener, J.F., Powell, D., Tregear, G.W., and Potts Jr., J.T. "Parathyroid Hormone in Human Plasma: Immunochemical Characterization and Biological Implications." **Journal of Clinical Investigations** 51:3163-3172, 1972.
- Freitag, J., Martin, K.J., Hruska, K.A., Anderson, C., Conrades, M., Ladenson, J., Klahr, S. and Slatopolsky, E. "Impaired Parathyroid Hormone Metabolism in Patients with Chronic Renal Failure." **New England Medical Journal of Medicine** 298:29-32, 1978.
- Potts Jr., J.T., Segre, G.V. and Endres, D.B. "Current Clinical Concepts: Assessment of Parathyroid Function with an N-Terminal Specific Radioimmunoassay for Intact Parathyroid Hormone." **Nichols Institute Reference Laboratories**, 1983.
- Goltzman, D., Henderson, B., and Loveridge, N. "Cytochemical Bioassay of Parathyroid Hormone: Characteristics of the Assay and Analysis of Circulating Hormonal Forms." **Journal of Clinical Investigations** 65:1309, 1980.
- Lafferty, F.W. " Pseudohyperparathyroidism." **Medicine** 45:247, 1966.
- Endres, D., Brickman, A., Goodman, W., Maloney, D., and Sherrard, D. "N-Terminal PTH Radioimmunoassays in Assessment of Renal Osteodystrophy." **Kidney International** 21:132, 1982.
- Broadus, A.E., Mahaffey, J.E., Bartter, F.C., and Neer, P.M. "Nephrogenous Cyclic Adenosine Monophosphate as a Parathyroid Function Test." **Journal of Clinical Investigations** 60:771, 1977.
- Berson, S.A., Yalow, R.S., Bauman, A., Rothchild, M.A. and Newerly, K. **Journal of Clinical Investigations** 35:170, 1956.
- Rodbard, D., Rayford, P.L., Cooper, J.A. and Ross, G.T. **Journal of Clinical Endocrinology Metab.** 28:1412, 1968.
- Segre, G.V. Niall, H.D., Habener, J.F., and Potts Jr., J. T. **American Journal of Medicine** 56:774.
- Flueck, J., Edis, A., McMahon, J. and Arnaud, C. "Proceedings of the 58th American Meeting of the Endocrine Society." June 1976.
- Silverman, R. and Yalow, R.S. **Journal of Clinical Investigations** 52:1958, 1973.
- Segre, G.V., Niall, H.D., Sauer, R.T. and Potts Jr., J.T. **Biochemistry** 16:2417, 1977.
- Canterbury, J.M., Bricker, L.A., Levy, G.S., Kozlovskis, et. al. **Journal of Clinical Investigations** 55:1245, 1975.
- Mallette, L.E., Tuma, S.N., Berger, R.E. and Kirkland, J.L. "Radioimmunoassay for the Middle Region of Human Parathyroid Hormone Using a Homologous Antiserum with a Carboxyl-terminal Fragment of Bovine Parathyroid Hormone as Radioligand." **Journal of Clinical Endocrinology Metab.** 54:1017, 1982.
- Roos, B.A., Lindall, A.W., Aron, J.W., et al. "Detection and Characterization of Small Mid-

- Region Parathyroid Hormone Fragments in Normal and Hyperparathyroid Glands and Sera by Immuno-Extraction and Region Specific Radioimmunoassays." **Journal of Clinical Endocrinology Metab.** 53:709, 1981.
23. Gallagher, J.C., Riggs, B.L., Jerpbak, C.M. and Arnaud, C.D. "The Effect of Age on Serum Immunoreactive Parathyroid Hormone in Normal and Osteoporotic Women." **Journal Of Laboratory Clinical Medicine** 95:373, 1980.
  24. Mallette, L.E. "Use of Homologous Antisera for Radioimmunoassay of Human Parathyroid Hormone." **Ligand Review** 1:18, 1979.
  25. Dambacher, M.A., Fischer, J.A., Hunziker, W.H. et al. "Distribution of Circulating Immunoreactive Components of Parathyroid Hormone in Normal Subjects and in Patients with Primary and Secondary Hyperparathyroidism: The Role of the Kidney and of the Serum Calcium Concentration." **Clinical Science** 57:435, 1979.
  26. Wood, W.G., Butz, R., Casaretto, M., et al. "Preliminary Results on the Use of an Anti-serum to Human Parathyrin in a Homologous Radioimmunoassay." **Journal of Clinical Chemical Biochemistry** 18:789, 1980.
  27. Kao, P.C., Jiang, N.S., Klee, G.G., and Purnell, D.C. "Development and Validation of a New Radioimmunoassay for Parathyrin (PTH)." **Clinical Chemistry** 28:69, 1982.
  28. Travis, J.C. (ed.) "Clinical Radioimmunoassay." **State-of-the-Art Scientific Newsletter, Inc.**, Anaheim, CA 92803, 1980.
  29. Rodbard, D., and Hutt, D. "Statistical Analysis of Radioimmuno-assays and Immunoradiometric (labeled antibody) Assays." **Assays, Radioimmunoassays and Related Procedures in Medicine**, Vol. 1 Vienna: International Atomic Energy Agency, Vienna, 1974.
  30. Nussbaum, S.R., Zahradnik, R.J., Lavigne, J.R., Brennan, G.L., Nozawa-Ung, K., Kim, L.Y., Kentmann, H.T., Wang, C.A., Potts Jr., J.T. and Segre, G.V. "Highly Sensitive Two-Site Immunoradiometric Assay of Parathyrin and Its Clinical Utility in Evaluating Patients with Hypercalcemia." **Clinical Chemistry** Vol. 33, No. 8, 1364-1367, 1988.
  31. Lepage R., Roy L., Brossard J.H., Rousseau L., Dorais C., Lazure C., D'Amour P. "Non-(1-84) Circulating Parathyroid Hormone (PTH) Fragment Interferes Significantly with Intact PTH Commercial Assay Measurements in Uremic Samples." **Clinical Chemistry** Vol. 44, No. 4, 805-809, 1998.
  32. Gao, P., Scheibel, S., D'Amour, P., Cantor, T.L.. "Measuring the Biologically Active or Authentic Whole Parathyroid Hormone (PTH) with a Novel Immunoradiometric Assay Without Cross-reaction to the PTH(7-84) Fragment." **Journal of Bone and Mineral Research** 14:S446, 1999.
  33. Brossard, J.H., Lepage, R., Gao, P., Cantor, T., Rousseau, L., D'Amour, P. "New Commercial Whole-PTH Assay Free of Interference by Non-(1-84) Parathyroid Hormone (PTH) Fragments in Uremic Samples." **Journal of Bone and Mineral Research** 14:S444, 1999.
  34. Slatopolsky, E., Finch, J.L., Martin, D., Sicard, G., Gao, P., Cantor, T. "A Novel Mechanism for Skeletal Resistance in Uremia." **Journal of American Society of Nephrology** 10:625A, 1999.
  35. John, M.R., Goodman, W.G., Gao, P., Cantor, T.L., Salusky, I.B., Jueppner, H. "A Novel Immunoradiometric Assay Detects Full-length Human PTH but Not Amino-terminally Truncated Fragments: Implication for PTH Measurements in Renal Failure." **The Journal of Clinical Endocrinology & Metabolism** 84:4287, 1999.
  36. Brossard J.H., Lepage R., Roy L., Rousseau L., Dorais C., Lazure C., D'Amour P. "Influence of Glomerular Filtration Rate on Non-(1-84) Parathyroid Hormone (PTH) Detected by Intact PTH Assays." **Clinical Chemistry** Vol. 46:5, 697-703, 2000.



**Scantibodies Laboratory, Inc.**

9336 Abraham Way

Santee, CA 92071, USA

Tel: (619) 258-9300

Fax: (619) 258-9366



Mandataire Europe:

**Laboratoire Scantibodies France**

12 rue de Normandie

91140 Villebon sur Yvette, FRANCE

Tel: 33-1 60 10 59 44

Fax: 33-1 60 10 76 41